ELMI

LABORATORY EQUIPMENT SINCE 1991

CE

ELMI SIA · RIGA · LATVIA

EU DECLARATION OF CONFORMITY

LABORATORY CENTRIFUGES

The undersigned ELMI SIA company declares, that the following instrument(s) conforms the below mentioned directives and standards:

MANUFACTURER	**ELMI SIA** Bukultu street 7b, Riga LV-1005, Latvia Phones: (+371) 67558 743, (+371) 67551 934
EQUIPMENT NAME	CM-50, CM-50M, CM-50MP, CM-7S Plus, CM-7S, CM-8S, CM-70M.07, CM-70M.09, CM-6M, CM-6MT, CM-70
EQUIPMENT TYPE	Laboratory centrifuge
SERIAL NUMBER	8 digits styled XX Y ZZZZ T , where XX is year of production, Y — quarter of production, ZZZZ – batch number, T – device modification
DIRECTIVES	EMC Directive 2014/30/EU RoHS 2011/65/EU Regulation (EU) 2017/746 Low Voltage Directive 2014/35/EU WEEE 2012/19/EU
APPLIED STANDARDS	EN 61326-1: 2013 — Electrical equipment for measurement, control and laboratory use EMC requirements. General requirements. EN 61010-1: 2010 — Safety requirements for electrical equipment for measurement, control and laboratory use. General requirements. EN 61010-2-020: 2016 — Particular requirements for laboratory centrifuges.

Vitalijs Mironovs
Technical Director

30.04.2025 DATE

Maksims Sviridovs
Manufacturing Director

30.04.2025 DATE