ELMI

LABORATORY EQUIPMENT SINCE 1991

CE

ELMI SIA · RIGA · LATVIA

DECLARATION OF CONFORMITY

ORBITAL SHAKERS

The undersigned ELMI SIA company declares, that the following instrument(s) conforms the bellow mentioned directives and standards:

MANUFACTURER	**ELMI SIA** Krīvu street 11-154, Riga LV-1006, Latvia Phones: (+371) 67558 743, (+371) 67551 934
EQUIPMENT NAME	S3, S-3M.A02, S-3M.A10, S-3M.A20, S-3L.A10, S-3L.A20, S-3.02M.A10, S-3.02M.A20, S-3.02L.A10, S-3.02L.A20, S-4
EQUIPMENT TYPE	Orbital Shaker
DIRECTIVES	EMC Directive 2004/108/EC Low Voltage Directive 2006/95/EC RoHS 2011/65/EC WEEE 2002/96/EC & 2012/19/EU In vitro diagnostic medical devices directive 98/79/EC
APPLIED STANDARDS	EN 61326-1: 2013 — Electrical equipment for measurement, control and laboratory use EMC requirements. General requirements. EN 61010-1: 2010 — Safety requirements for electrical equipment for measurement, control and laboratory use. General requirements. EN 61010-2-051: 2015 — Particular requirements for laboratory equipment for mixing and stirring.

Vitalijs Mironovs
Technical Director

30.04.2019 DATE

Maksims Sviridovs
Manufacturing Director

30.04.2019 DATE